

# HYGRO THERMOMETER MODEL: 820





## **FEATURES:**

Highly stable humidity measurement ±2.5%. Processing %RH and temperature simultaneously. Two in one wide range thermometer and humidity function. Wet/Dry bulb, dew point and %RH. Easy switching between K/NTC/K-DP. Max/Min/AVG/REL/HOLD function. APO function.

## **SPECIFICATIONS**

Display: 3 ½ digit liquid crystal display (LCD) with a maximum reading of 1999.

Polarity: Automatic, positive implied, negative polarity indication. Zero: Automatic.

Low battery indication: The "" is displayed when the battery voltage drops below the operating level.

Operating environment: 0°C to 50°C at<75% R.H.

Storage temperature:-20°C to 60°C, 0 to 80% R.H with battery removed from meter.

Accuracy: Stated accuracy at 23°C±5°C, <75% R.H.

Battery: 1.5Vx4pcs AAA size.

#### TEMPERATURE RANGE

Range: -50°C to 1000°C, -50°F to 1832°F Accuracy: ±(0.1% rdg+1°C) on 50°C to 1000°C ±(0.1% rdg+2°F) on -50°F to 1832°F

#### **TEMPERATURE**

Sensor: Thermistor temperature sensor Range: 0°C to 60°C, 32°F to 140°F Resolution: 0.1°C/°F Accuracy:

±2°C on 0°C to 10°C

±0.5°C on 10°C to 45°C

±2°C on 45°C to 60°C

±4°F on 32°F to 50°F

±1°F on 50°F to 113°F

# RELATIVE HUMIDITY

Sensor: Capacitive humidity sensor

Range: 0% to 100% RH

±4°F on 113°F to 140°F

Accuracy:

±2.5% at 25°C(77°F), 10% to 90% RH

±5% at 25°C(77°F), 0% to 10% RH, 90% to 100% RH

Sensor Response Time for 90% of Total Range:

60sec typical.

Sensor Hysteresis(excursion of 10% to 90% to 10% RH):

±1%RH typical.

**CHY FIREMATE CO., LTD.** 

NO.3, SHENGLI 1<sup>st</sup> ST., XINTIAN VILLAGE, RENDE DISTRICT, TAINAN CITY, TAIWAN R.O.C.

TEL: 886-6-2794811 FAX: 886-6-2700227

E-mail: <a href="mailto:chy@chy-meter.com">chy@chy-meter.com</a>