


[REC]
[OFF]

Start the recording.
Stop the recording.

Open\Display\Ghartgraph



Button	Description
[Exit]	To exit 792R/792RT application.

III Setup, Start & Exit on Software

Quick Start at the 792R/792RT

1. Install the RS-232 cable between PC's Com port and 792R's/792RT's RS-232 port.
Please make sure:
 - The RS232 cable connector (to meter) faced up with panel.
 - Physical Com port number has to be the same as setup in software.
2. Turn the 792R/792RT on.

At the PC



1. Open the 792R/792RT software by double-clicking on the icon SETUP.exe.
2. Please double-clicking 792R/792RT on the file C:\792R\ to start the 792R application.

C. Quick Exit

1. Click Exit to exit 792R/792RT application.

Operation Manual

MODEL 792/792R/792RT




TABLE OF CONTENTS	PAGE
1. Introduction	1
2. General	1
3. Specifications.....	1
4. Front Panel Description	2
4.1 Push Button functions	2
4.1.1 “HLD” Button	2
4.1.2 “MAX/MIN” Button	2
4.1.3 “pH/mV” Button	2
4.1.4 “MEAS” Button	3
4.2 Switch Functions.....	3
4.2.1 “°C/°F” Switch.....	3
4.2.2 “CAL” Switch.....	3
5. pH Temperature Compensation.....	3
5.1 Temperature Compensation Mode	3
5.1.1 Automatic Temperature Compensation	3
5.1.2 Manual Temperature Compensation	3
6. Calibrating Procedure.....	3
6.1 pH Calibration With Automatic Temperature Compensation(ATC).....	3
6.2 pH Calibration With Manual Temperature Compensation(MTC)	4
6.3 Temperature Calibration	6
7. Measuring Procedure	7
7.1 pH Measurement	7
7.2 mV Measurement	7
8. pH Electrode Maintenance	7
8.1 Cleaning the pH Bulb	7
8.1.1 Protein Contamination	7
8.1.2 Oil Contamination.....	7
8.2 Recondition the pH Bulb.....	7
9. Trouble Shooting	8
10. Battery Replacement	8
11. Cleaning.....	8
I. Install the 792R/792RT Software.....	9
II. 792R/792RT Software Operate manual.....	9
III. Setup, Start & Exit on Software.....	10

I Install the 792R/792RT Software

1. Run setup.exe on your 792R/792RT software.
2. Change the path if necessary or choose “Finish” button to install program directly.

Note:

1. Check the 792R/792RT software for virus before installation.
2. System required: 80486-33 with 16MB RAM or better.


II 792R/792RT Software Operate manual



File

- Open File To Excel To open excel file of REC data and save data.
- Print Window Print Main Window

Open/Display/List




Button	Description
[FILE NAME]	1. DEFAULT-saving: the file name saved with time.xls Such as: C:\792R\data\06161133.xls. This file was saved at 11:33 am. June 16.
[INTV]	2. User-saving: the file name saved by user.
[LIST]	Enter the number of second between recording input reading.
[CLR]	Enter the number of message history then view the message by shifting bar. Clear the message window.

9. Trouble Shooting

- ? Power on but no display: Check the battery are in place and making good contact or correct polarity, replace a new battery.
- ? Unstable reading: Clean the probe and recalibrate or make sure sample entirely covers the sensor, or replace a new probe and re-calibrate if you find a broken probe.
- ? Slow response: Clean probe by immersing the electrode in tap water for 10-15 minutes, then thoroughly rinse with distilled water or use a general purpose electrode cleaner.
- ? LCD show E2 at ATC Range: The temperature of solution over 100°C (212°F).
- ? LCD show E3 at ATC Range: The temperature of solution under 0°C (32°F).

10. Battery Replacement

1. Power is supplied by three 1.5V(AAA size) batteries.
2. The “” appears on the LCD display when replacement is needed. To replace the batteries, remove the screw from the back of the meter and lift off the battery cover.
3. Install three 1.5V batteries.
4. When not use for long time remove battery.
5. Don't keep in place with high Temp. or high humidity.


11. Cleaning

Periodically wipe the case with a damp cloth and detergent, do not use abrasives of solvents.

1. Introduction

- This instrument is a portable easy to use 3 ½ digit, compact-sized pH/mV and temperature meter.
- Dual display pH (mV) & Temperature.
- Auto power off function to extend battery life.
- Automatic Temperature Compensation and Manual Compensation.
- RS 232 Interface with Windows Software.(792R/792RT)
- Software download link:
http://WWW.chy-meter.com//data/outer_web/1492-00792R/1492-00792R-101.zip

2. General

Low battery indication: The “” is displayed when the battery voltage drops below the operating level.

Accuracy: Stated accuracy at 23°C±5°C (79°F±9°F), < 70% R.H.

Operating environment:

Meter: 0°C to 50°C (32°F to 122°F) at < 70% R.H.

pH electrode: 0°C to 80°C (32°F to 176°F).

Storage environment: -20°C to 60°C (-4°F to 140°F) at < 80% R.H. with battery and pH electrode removed form meter.

Measurement rate: 1 time per second, nominal.

EMC-instrument unspecified for use in EMC field ≥ 0.5V/m.

Battery: 3 pcs 1.5V (AAA size) UM-4 R03.

Battery life: 100 hours typical with carbon zinc battery.

Dimensions: 147mm(H) x 48mm(W) x 30mm(D).

Weight: 180g.

Accessories:

792: Instruction manual, Batteries, pH4.01, pH7.00 and pH10.01 buffer solution, pH electrode, screwdriver.

792R: Instruction manual, Batteries, RS232 cable, pH4.01, pH7.00 and pH10.01 buffer solution, pH electrode, screwdriver.

792RT: Instruction manual, Batteries, Temperature probe, screwdriver.

The screwdriver: It is used to select temperature in °C or °F and CAL position.

Option Accessories: pH electrode, pH4.01, pH7.00 and pH10.01 buffer solution, Carrying case.

3. Specifications

Accuracy stated at ambient temperature 18 to 28°C(65 to 82°F), <70% R.H.

Temperature Coefficient: 0.1 times the applicable accuracy per °C from 0°C to 18°C and 28°C to 50°C(32°F to 64°F and 82°F to 122°F)


Measurement	Range	Resolution	Accuracy
pH	0.00 to 14.00	0.01pH	±0.03pH*
mV	-1000 to 1000mV(Auto)	0.1mV/1mV	±(0.5% rdg+2dgts)
Temp(°C)	0 to 80°C	0.1°C	±1°C
Temp(°F)	32 to 176°F	0.1°F	±2°F

*Not include pH electrode error

Input impedance: 10¹²Ω

Note: Probe can't be interchanged.

4. Front Panel Description



1. LCD: Measured values, unit, symbols and decimal points are displayed.
2. Input Socket: BNC connector for pH and mV.
3. Input Socket: Earphone (3.5mm) jack for temperature plug.
4. Input Socket: Earphone (2.5mm) jack for RS232 cable.(792R/792RT)
5. Push Buttons: Button for controlling meter.
6. Push Button: Turn on/off meter.
7. CAL Switch: Slide the switch to right to enter calibration mode.
8. °C/°F Switch: Slide the switch °C/°F to select the temperature unit.

4.1 Push Button functions

4.1.1 “HLD” Button:

Press the “HLD” button to enter the data hold mode, the “HLD” annunciator is displayed at the top-right of display. When data hold mode is selected, the pH meter held the present readings and stops all further measurements. Press the “HLD” button again to cancel data hold mode, causing pH meter to resume taking measurements.

4.1.2 “MAX/MIN” Button:

Press the “MAX/MIN” button to enter the MAX, MIN, MAX-MIN, AVG recording mode. (displays the Maximum reading, Minimum reading, “MAX-MIN” reading and “AVG” reading in record mode)

In this mode, press “HLD” button to stop recording, all values are frozen, press “HLD” button again to restart recording. In this mode, the APO function and other buttons are disabled, excluding “HLD” and Back-light buttons. Press and hold down the “MAX/MIN” button for more than 2 seconds to exit the MAX/MIN function.

4.1.3 “pH/mV” Button:

It built in mV measuring function, letting you make ion-selective, ORP(oxidation-reduction potential), and other precise mV measurements. Pressing “pH/mV” button to change measurement range between “pH” and “mV”.

7. Measuring Procedure

7.1 pH Measurement

Calibrate the instruments and pH electrode before measuring.

1. Connect the combination pH electrode to the BNC socket.
2. Power on the instrument by pressing the “MEAS” button.
3. If the operation is under the “ATC” then please refer 5.1.1 Automatic Temperature compensation mode.
4. If the operation is under the “MTC”, (take the 3.5 mm temperature plug away) then please refer 5.1.2 Manual Temperature compensation mode.
5. Place the electrode into the measured solution, the instrument will display the pH value.
6. After making the measurement, please rinse the electrode with distilled water.

7.2 mV Measurement

The instrument builds in mV measuring function letting you make ORP or other precise mV measurements.

1. Power on the instrument by pressing the “MEAS” button and pressing pH/mV button to change unit between “pH” and “mV”.
2. The meter will show the mV value on the display.

8. pH Electrode Maintenance

The proper way of using and protecting the electrode, it will prolong the life of the glass membrane. If your pH electrode is exhibiting by slow response, continuous drift, or erratic readings, follow the procedures listed below.

8.1 Cleaning the pH Bulb

8.1.1 Protein Contamination:

Soak the electrode bulb/tip in a 10% solution of pepsin for 30 minutes. Rinse with deionized water and soak the electrode in pH7.00 buffer for two hours before using.

8.1.2 Oil Contamination:

Wash the electrode with a 50% water-acetone solution. Do not soak the electrode in the acetone solution, or it will deteriorate the bottom seals of the plastic electrode, Rinse with deionized water and soak the electrode in pH7.00 buffer for two hours before using.

8.2. Recondition the pH Bulb

Only resort to this procedure if the preceding maintenance and cleaning procedures fail to restore acceptable electrode performance. Rinse immediately with deionized water and soak in pH7.00 buffer for two hours before using.

Caution

To prevent permanent damage, care should be taken to prevent liquid permeating the pH meter. Meanwhile, the batteries should be taken out if user will not use the meter for a long period. Also, to choose the fitted pH electrode is required. Please always keep the pH glass bulb wet by 4M KCL. Always rinse the pH electrode and reference junction in de-ionized water before next use. Never touch or rub glass bulb for lasting pH electrode life.

6.3 Temperature Calibration:

1. Select the function in mV and connect the temperature plug.
2. Sliding the “CAL” switch to right to enter calibration procedure.
3. Put pH electrode into a standard which is 25.0±10°C (or 77.0±18°F) and the temperature is stable, if the temperature is inaccurate, LCD will show E5, please check value of temperature.(Fig-13)
4. Waiting for the reading to be stable. (Fig-14)
5. Press and hold down “SET” button for more than 2 seconds to enter the temperature calibration, and the sub (Temp.) display blanking, (Fig-15)




Fig-13




Fig-14




Fig-15

6. The preset value is 25.0°C (or 77.0°F) or user can press “▲” or “▼” button which are printed in white sign to increase or decrease the determined value in LCD. Press and hold down “SET” button for more than 2 seconds to complete 25.0°C (or 77.0°F) calibration, and LCD will show SA. (Fig-16, Fig-17)




Fig-16




Fig-17

7. Put pH electrode into a standard which is 70.0±10°C (or 158.0±18°F) and the temperature is stabled, repeat steps 3 to 6.
8. If you complete steps 1 to 7, the LCD will show CAL(main display) and End(sub display). (Fig-18)



Fig-18

9. Sliding the “CAL” switch to left to leave calibration procedure.

4.1.4 “MEAS” Button:

It can turn on or off the meter by pressing “MEAS” button, power is automatically turn off if no operation for a period of 10 min.

4.2 Switch Functions

4.2.1 “°C/°F” Switch:

It can select the temperature unit by sliding the switch. If intend to measure the “°C”, then slide the switch to left. If intend to measure the “°F”, then slide the switch to right.

4.2.2 “CAL” Switch:

It can enter the calibration procedure by sliding the switch to right.

5. pH Temperature Compensation

Enable the meter to read solutions at various temperatures, the meter will make the correct electrode’s temperature dependency to measure the pH value. The compensation may be manual with two buttons adjustment on the meter, or it may be automatic with pH electrode’s temperature sensor immersed in test solution.

5.1 Temperature Compensation Mode:

5.1.1 Automatic Temperature Compensation:

Connecting the 3.5mm temperature plug into earphone and the pH function will be temperature compensated automatically.

5.1.2 Manual Temperature Compensation:

Take the 3.5mm temperature plug away then press and hold down the “SET” button which are printed in white word on the overlay for more then 2 seconds to enter temperature compensation mode between 0.0°C and 100.0°C (32.0°F to 212°F). The preset temperature is 25.0°C (77.0°F), press “▲” or “▼” which are printed in white sign to increase or decrease the compensation temperature. Pressing and hold down “SET” button more than 2 seconds to complete the temperature.

6. Calibration Procedure

6.1 pH Calibration With Automatic Temperature Compensation(ATC):

1. Connect the 3.5mm temperature plug into earphone. Select the function in pH(ATC)/°C. (or °F) Slide the switch °C/°F to select the temperature unit.
2. Sliding the “CAL” switch to right to enter calibration procedure.
3. Put pH electrode into buffer solution pH4.01 and agitate the buffer solution. If the solution is inaccurate, LCD will show E4, please check value of buffer solution. (Fig-1)
4. Wait for the reading to be stable. (Fig-2)




Fig-1




Fig-2

- Press and hold down “SET” button for more than 2 seconds to enter the pH4.01 mode, and the main (pH) display blanks. (Fig-3)
- The preset value is pH4.01 and user can press “▲” or “▼” button to enter determine value in LCD. Press and hold down “SET” button for more than 2 seconds to complete the pH4.01 calibration, and LCD will show SA. (Fig-4, Fig-1)




Fig-3




Fig-4

- Rinse electrode and blot with lint-free tissue.
- Put pH electrode into buffer solution pH7.00, repeat steps 3 to 7.
- Put pH electrode into buffer solution pH10.01, repeat steps 3 to 7.
- If you complete steps 1 to 9, the LCD will show CAL(main display) and End(sub display). (Fig-5)
- Sliding the “CAL” switch to left to leave calibration procedure.




Fig-5

6.2 pH Calibration With Manual Temperature Compensation(MTC):

- Take the 3.5mm temperature plug away. Select the function in pH(MTC)/°C. (or °F)
- Sliding the “CAL” switch to right to enter calibration procedure.
- Put pH electrode into buffer solution pH4.01 and agitate the buffer solution, if the solution is inaccurate, LCD will show E4, please check value of buffer solution. (Fig-6)
- Waiting for the reading to be stable. (Fig-7)




Fig-6




Fig-7

- Press and hold down “SET” button for more than 2 seconds to enter the pH4.01 calibration procedure, and the sub(Temp.) display blanking. (Fig-8)




Fig-8

- The preset value is 25.0°C (77.0°F) and user can press “▲” or “▼” buttons which are printed in white sign to increase or decrease determined value in LCD. Press and hold down “SET” button for more than 2 seconds to complete the temperature of pH4.01 calibration, and LCD will show SA(Fig-9), then the main display blanking. (Fig-10)




Fig-9




Fig-10

- The preset value is pH 4.01 and user can press “▲” or “▼” buttons which are printed in white sign to increase or decrease determined value in LCD. Press and hold down “SET” button for more than 2 seconds to complete pH4.01 calibration, and LCD will show SA. (Fig-9, Fig-11)




Fig-11

- Rinse electrode and blot with lint-free tissue.
- Put pH electrode into buffer solution pH7.00, repeat steps 3 to 8.
- Put pH electrode into buffer solution pH10.01, repeat steps 3 to 8.
- If you complete steps 1 to 10, the LCD will show CAL(main display) and End(sub display). (Fig-12)




Fig-12

- Sliding the “CAL” switch to left to leave calibration procedure.